

**STABILITY STATEMENT for
TaqPath™ ProAmp™ Master Mix and TaqPath™ ProAmp™ Multiplex Master Mix**

Thermo Fisher Scientific confirms that the stability for TaqPath™ ProAmp™ and TaqPath™ ProAmp™ Multiplex Master Mix was evaluated by performing accelerated, in-use, transportation and long-term stability studies (SS). The expiration date was determined based on the performance characteristics of product stored under specified experimental conditions per EN ISO 23640:2015 In vitro diagnostic medical devices - Evaluation of stability of in vitro diagnostic reagents. Exact expiry date is printed on the product label and reflected within the lot specific Certificate of Analysis (CoA). **The products remain stable under the described SS conditions (accelerated, in-use, transportation and long-term SS) and maintain their performance for at least 12 months post manufacturing when held at 5±3°C.**

Catalog SKUs in scope of this Stability Statement include:

TaqPath™ ProAmp™ Master Mix	TaqPath™ ProAmp™ Multiplex Master Mix	Amount	Storage
with ROX™	with MUSTANG PURPLE™		
A30865	A30868	1 x 1 mL	5±3°C
A30866	A30869	1 x 10 mL	
A30871	A30873	2 x 10 mL	
A30867	A30870	1 x 50 mL	
A30872	A30874	2 x 50 mL	

Authorized by:

Senior Quality Manager

Ana Mačernienė



Electronically signed by:
Ana Mačernienė
Reason: Approver of the
QAP document
Date: Jun 3, 2022 12:45
GMT+3