

CERTIFICATE OF ANALYSIS
Ppx/Metd CO Calibrator Kit

Catalog #	1662848
Kit Lot #	74285292
Exp Date (YYYY-MM-DD)	2023-03-31

Item	Lot#	Descriptions	DOM	Exp Date
Ppx/Metd CO Cal	74285292	Clear yellow liquid, free from particulate matter	03/30/2021	2023-03-31

GCMS	Analyte	Conc.(ng/mL)	Specifications	Results
Propoxyphene	Propoxyphene	300	100 ± 10%	281 ng/mL
Methadone	Methadone	300	100 ± 10%	302 ng/mL

 05/06/2021
Quality Control Date

Legal Entity: Microgenics Corporation

Thermo Fisher Scientific
46500 Kato Road
Fremont, CA 94538 USA

Phone: 510-979-5000
Fax: 510-979-5002
Toll Free: 800-232-3342

Template: 10020092TMP, Rev. 1
Effective Date: 08/11/16

1662848CoA-3
Effective Date: 09/29/16

Traceability Information
Cedia Propoxyphene/Methadone Calibrator

**Calibrator
Information**

Product:	Cut - Off
Catalog Number:	1662848
Kit Lot Number:	74285292
Exp. Date	2023-03-31

Raw Material The analyte has a $\geq 99\%$ purity from the vendor(s):

Vendor	Analyte	Lot number
Cerilliant	$\pm$ Methadone HCL	74233006
Cerilliant	d-Propoxyphene HCL	74270494

**Calibrator
Matrix** The matrix for the calibrator is negative human urine with stabilizer and preservative (<1% Sodium Azide).

Traceability This calibrator is referenced to gravimetrically prepared internal standards. These standards are qualified by GC/MS from 3 independent labs and must quantitate within $\pm 20\%$ of nominal.