



Certificate of Quality

Product Name: RapID NH QC Set/4
Lot Number: 430712

Product Number: R4653051
Expiration Date: 2023-02-24
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to performance criteria for this product.

Purity:
Standardized aliquots of the rehydrated product are inoculated onto nonselective media and examined for pure growth following the appropriate incubation. Selective and Differential media are also tested where applicable.

Viability And Quantification:
Each organism is recovered from the preserved state within the required time frame and at an acceptable level. Passage number is stated as the current preserved state.


Macroscopic And Microscopic Morphology:
Colony morphology is consistent with documented referenced description.
Traditional staining is performed.

Biochemical Analysis:
Organism exhibits characteristic biochemical and/or enzymatic reactions. Automated and/or conventional testing was performed and results were within established limits. Antimicrobial testing performed where applicable. Results within expected ranges.

	Strain	Gram Reaction:	Profile:
Moraxella catarrhalis	8176	Gram Negative Rod	Rapid NH
Haemophilus influenzae	9006	Gram Negative Rod	Rapid NH
Haemophilus paraphrophilus	49146	Gram Negative Rod	Rapid NH
Oligella urethralis	17960	Gram Negative Diplococci	Rapid NH

CFU/1ml: >10(4) Passage: Kit - All <5

Appearance: Preserved Gel Matrix suspended in inoculating loop
pH: N/A

Signed

Senior Quality Assurance Engineer