Release Notes
Qubit™ 4 Fluorometer Firmware v2.15

Contents
- OVERVIEW
- FEATURES
- DOWNLOAD INSTRUCTIONS
- UPGRADE INSTRUCTIONS
- LEGAL STATEMENTS

Overview
Firmware v2.15 for the Qubit™ 4 Fluorometer

This firmware release is for Qubit™ 4 Fluorometer only:

Q33226   Qubit™ 4 Fluorometer

New Features/Changes of version 2.15
- Add support for Qubit™ RNA XR
- SB-327 compliance

Downloading the Firmware File
- Download the attached file onto a USB memory stick (shown in .pak format)
- Save/Transfer the .pak file (not in any folder) on your USB stick
- Make sure that no other .pak files are present on the memory stick

Updating your Qubit™ instrument

1. Power on the instrument and log-in as a user
2. Insert the USB memory stick containing the firmware update into one of the USB ports in the Qubit ™ instrument.
3. On the Home screen, touch 'Settings'.
4. In the Settings menu, touch 'Update Software'.
5. If you have downloaded the firmware properly, the instrument should locate the file.
6. Press ''Upgrade' at the bottom of the screen to confirm that you want to upgrade.

IMPORTANT! Ensure that you do not turn off the instrument when the upgrade is in progress.

7. When the upgrade process is complete, the instrument will restart automatically. When it has rebooted, you can remove your USB stick.

LEGAL STATEMENTS

Important licensing information:

These products may be covered by one or more Limited Use Label Licenses. By use of these products, you accept the terms and conditions of all applicable Limited Use Label Licenses.

The trademarks mentioned herein are the property of Life Technologies Corporation and/or its affiliate(s) or their respective owners.

For Research Use Only. Not for use in diagnostic procedures.