

Catalog Number B34956
Product Name BacLight™ RedoxSensor™ CTC Vitality Kit *for flow cytometry and microscopy*
Lot Number 2263420

	LOT DATA	SPECIFICATION
ABSORPTION Optical Density of Component B ¹	0.69 at 486 nm	0.68 ± 0.18 at 485 ± 3 nm
ABSORPTION Optical Density of Component C ²	0.90 at 348 nm	0.89 ± 0.09 at 348 ± 3 nm

1. Buffer: Methanol (50x dilution)
2. Buffer: Methanol (500x dilution)



Rachel Smith, Quality Assurance Manager
 10-May-2021

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.