

Catalog Number L34952
Product Name LIVE/DEAD® FungaLight™ Yeast Viability Kit *for flow cytometry*
Lot Number 2532514

	LOT DATA	SPECIFICATION
ABSORPTION Optical Density of Component A ¹	0.71 at 474 nm	0.75 ± 0.08 at 474 ± 3 nm
ABSORPTION Optical Density of Component B ²	0.62 at 494 nm	0.59 ± 0.06 at 493 ± 3 nm

1. Buffer: phosphate buffered saline, pH 7.2 (301x dilution)

2. Buffer: phosphate buffered saline, pH 7.2 (201x dilution)



Aaron Rider, Quality Control Supervisor
 27-Sep-2022

Life Technologies Corporation certifies on the date above that this is an accurate record of the analysis of the subject lot, and that the data conform to the specifications in effect for this product at the time of analysis.