

Certificate of Analysis

Product Name: Blood Bank Saline (10L)
Product Numbers: 062-125
Lot Number: 343341

Production Date: 18-AUG-18
Expiration Date: 2020-02-29

Specification	Specification Limit	Result
Microbial Content	≤ 1 CFU/L	0
pH	6.0 - 7.5	6.3
Direct Antiglobulin	Complies	Complies
Osmolality	275 - 295 mOsm	286
Isotonicity	No observable hemolysis with human red blood cells	Complies
Specificity	a) Normal red blood cells observed microscopically b) No red cell aggregates c) No positive anti-human globulin d) Normal morphology	Complies



Quality Representative

Date: 2018-08-24