

## CERTIFICATE OF ANALYSIS

### SMC Bromperidol Kit Calibrators

Microgenics Catalog # :	300008
Kit Lot # :	74492643
Kit Date of Manufacture:	12/08/2021
Exp. Date:	2023 12

Item	Lot #	DOM	Exp. Date	Fill Volume
Low Bromperidol Calibrator	74492687	12/08/2021	2023 12	7.5 ml
High Bromperidol Calibrator	74492694	12/08/2021	2023 12	5.0 ml

Master Lot # :	74238958 74398008
Test Date :	01/18/2022

Item	Assigned Value	Specification	Pass/Fail
Low Bromperidol Calibrator	X = 0.43 ng/mL	0 - 1.25 ng/mL	PASS
High Bromperidol Calibrator	X = 39.2 ng/mL	36.0 - 44.0 ng/mL	PASS

  
Quality Control

01/21/2022  
Date

300008CoA-4  
Effective Date: 08/04/16

300007CoA-4  
Effective Date: 08/04/16

Legal Entity: Microgenics Corporation

Thermo Fisher Scientific  
46500 Kato Road  
Fremont, CA 94538 USA

Phone: 510-979-5000  
Fax: 510-979-5002  
Toll Free: 800-232-3342

Template: 10020092TMP, Rev. 1  
Effective Date: 08/11/16