

Certificate of analysis

1,1,1,3,3,3-Hexafluoro-2-propanol, 99%

Stock Number: 15193

Lot Number: G08N10

Analysis

Assay	99+ % by VPC Column: DB-1 (100% -polydimethylsiloxane) Length: 30m; ID: 0.25mm; Film: 1.0 μ l Injection temp: 250°C Detector (FID) temp: 275°C Column temp: 40°C initial, hold for 4 min, then programmed at 10°C/min to 250°C, hold for 10 min Sample: 0.2 μ l neat
Appearance	Clear, colorless liquid

Certified by:



Quality Control

This document has been electronically generated and does not require a signature.