

# Certificate of analysis

**1,1,1,3,3,3-Hexafluoro-2-propanol, 99%**


**Stock Number: 15193**

**Lot Number: H11P05**

## Analysis

Assay	99+ % by VPC Column: DB-1 (100 % -polydimethylsiloxane) Length: 30 m; I.D.: 0.25mm; Film: 1.0 µl Injection temp: 250°C Detector (FID) temp: 275°C Column temp: 40°C initial, hold for 4 min, programmed at 10°C/min to 250°C, hold for 15 min Sample: 0.2 µl neat
Appearance	Clear, colorless liquid

Certified by:



Quality Control

This document has been electronically generated and does not require a signature.