

Tel 585-586-8800 Fax 585-899-7605 75 Panorama Creek Drive, Rochester, NY 14625 Product Certificate
Thermo Scientific
Nalgene and Nunc Products

Thermo Fisher Scientific hereby certifies that the product identified below is produced, inspected and found to be in compliance with product and quality specification requirements as documented in our ISO 13485:2003 Quality Management System (CERT-0040858 by QMI-SAI Global) in the USA.

RIDM Rent Robert Prescott Mgr. QA/RA

The following information represents Product Certification for: Item#: 362150-1384

Certificate issued: 09/14/2009

Description: COLOR CLSR,PP,38-430,WHITE Lot#: 7105836010

Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
CLOS,38/430,PP,WHT,NALGE	COMPONENT PART				
RESIN,PP,WHI,INJ	POLYPROPYLENE, WHITE, INJ.	N/A	PASSED	PASSED	N/A
RESIN,PP,INJ	POLYPROPYLENE, INJECTION	12204	PASSED	PASSED	177.1520(a)(1)(i), (c)1.1a(use conditions A-H)
COLOR,WHT,MULTI	COLORANT, WHITE	16513	PASSED	PASSED	177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28.184.1210
	CLOS,38/430,PP,WHT,NALGE RESIN,PP,WHI,INJ RESIN,PP,INJ	CLOS,38/430,PP,WHT,NALGE RESIN,PP,WHI,INJ RESIN,PP,INJ COMPONENT PART POLYPROPYLENE, WHITE, INJ. POLYPROPYLENE, INJECTION	CLOS,38/430,PP,WHT,NALGE COMPONENT PART RESIN,PP,WHI,INJ POLYPROPYLENE, WHITE, INJ. N/A RESIN,PP,INJ POLYPROPYLENE, INJECTION 12204	CLOS,38/430,PP,WHT,NALGE COMPONENT PART RESIN,PP,WHI,INJ POLYPROPYLENE, WHITE, INJ. N/A PASSED RESIN,PP,INJ POLYPROPYLENE, INJECTION 12204 PASSED	CLOS,38/430,PP,WHT,NALGE COMPONENT PART RESIN,PP,WHI,INJ POLYPROPYLENE, WHITE, INJ. N/A PASSED PASSED RESIN,PP,INJ POLYPROPYLENE, INJECTION 12204 PASSED PASSED

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.