

Tel 585-586-8800 Fax 585-899-7605 75 Panorama Creek Drive, Rochester, NY 14625

Product Certificate Thermo Scientific Nalgene and Nunc Products

Thermo Fisher Scientific hereby certifies that the product identified below is produced, inspected and found to be in compliance with product and quality specification requirements as documented in our ISO 13485:2003 Quality Management System (QMI-SAI Global File No. 1616978) in Monterrey, Mexico.

Lisa Adams
Mgr. Quality Engineering

The following information represents Product Certification for: Item#: 190-9920

Certificate issued: 11/14/2012

Description: SRNG FLTR BLK SFCA;25MM,.2UM]] Lot#: 1084048 Use Before: 11/28/2017 Manufactured: 11/01/2012

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-0065-86	LOWER, SYR/FLT, 25MM, ACR, NAT	COMPONENT PART				
8-0092-02	RESIN, A CR, TERPOLYMER, INJ	POLYMETHYL METHACRYLATE	BB-MF-2138	PASSED	PASSED	177.1010, 176.170
1-0065-92	UPPER,SYR/FLT,25MM,ACR,NAT	COMPONENT PART				
8-0092-02	RESIN, A CR, TERPOLYMER, INJ	POLYMETHYL METHACRYLATE	BB-MF-2138	PASSED	PASSED	177.1010, 176.170

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Product was Gamma Irradiation Sterilized. Product was dosimetric released per ANSI/AAMI/ISO 11137 guidelines. Product was determined to be non-pyrogenic at a level < 0.5 EU/ml as by "Guidelines on the Validation of the Limulus Amebocyte Lysate Test," as defined by the FDA (12/87), as an end product endotoxin test for human and animal parenteral drugs, biological products and medical devices.