

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485:2016 (BSI Certificate Number: FM 653694) in Monterrey Mexico.



Ana Martinez
Quality Manager

Certificate issued: 11/18/2019

The following information represents Product Certification for: Item#: **FB12566502**

Description: **FB FILTER, PES, 250ML, 50MM, 0.2UM**

Lot#: **1269904**

Use Before: **11/28/2024**

Manufactured: **11/08/2019**

| Part Number | Description | Common Name | DMF# | Cytotoxicity | USP Class VI | FDA Compliance - 21 CFR |
|-------------|--|-------------------------------|------|--------------|--------------|--|
| 1-0014-16 | COVER, 150mL/250mL, FILTER UNIT, FISHERB | COMPONENT PART | | | | |
| 8-0077-12 | RESIN,PS,INJ,IBM | POLYSTYRENE | 1623 | PASSED | PASSED | 177.1640 |
| 1-0014-91 | UPPER,FLT,250 ML, MF-75, PS | COMPONENT PART | | | | |
| 8-0077-12 | RESIN,PS,INJ,IBM | POLYSTYRENE | 1623 | PASSED | PASSED | 177.1640 |
| 1-0014-63 | LOWER,FLT,250ML,45MM NK,ANON | COMPONENT PART | | | | |
| 8-0077-12 | RESIN,PS,INJ,IBM | POLYSTYRENE | 1623 | PASSED | PASSED | 177.1640 |
| 1-1804-61 | CLOS,45/430,HDPE,BLU,ANON | COMPONENT PART | | | | |
| 8-0042-18P | RESIN,HDPE,BLU,INJ | COLOR MIX (RESIN, HDPE, BLUE) | N/A | PASSED | PASSED | N/A |
| 8-0042-01 | RESIN,HDPE,INJ | HIGH-DENSITY POLYETHYLENE | 1646 | PASSED | PASSED | 176.170(c), 177.1520(c)3.2a |
| 8-0099-74 | COLOR,BLUE,MULTI | COLORANT, BLUE, MULTI-RESIN | N/A | PASSED | PASSED | 177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210 |
| 1-0014-82 | COLLAR,FLT,45/50,MF75,HIPS,BLU | COMPONENT PART | | | | |
| 8-0077-18P | RESIN,PS,HIGH IMPACT,BLU,INJ | COLOR MIX | N/A | PASSED | PASSED | N/A |
| 8-0077-13 | RESIN,PS,HIGH IMPACT,INJ | POLYSTYRENE | 1623 | PASSED | PASSED | 177.1640 |
| 8-0099-74 | COLOR,BLUE,MULTI | COLORANT, BLUE, MULTI-RESIN | N/A | PASSED | PASSED | 177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210 |
| 1-0014-95 | ADP,FLT,Q/D,LDPE,WHT,MF-75 | COMPONENT PART | | | | |
| 8-0049-14P | RESIN,LDPE,WHT,INJ | COLOR MIX (LDPE, WHT, INJ) | N/A | N/A | N/A | N/A |
| 8-0049-01 | RESIN,LDPE,INJ | LOW-DENSITY POLYETHYLENE | 9522 | PASSED | PASSED | 177.1520(a)(2)(i) |

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485:2016 (BSI Certificate Number: FM 653694) in Monterrey Mexico.



Ana Martinez
Quality Manager

Certificate issued: **11/18/2019**

The following information represents Product Certification for: Item#: **FB12566502**

Description: **FB FILTER, PES, 250ML, 50MM, 0.2UM**

Lot#: **1269904**

Use Before: **11/28/2024**

Manufactured: **11/08/2019**

| Part Number | Description | Common Name | DMF# | Cytotoxicity | USP Class VI | FDA Compliance - 21 CFR |
|-------------|---------------|-----------------|------|--------------|--------------|--|
| 8-0099-26 | COLOR,WHT,OLF | COLORANT, WHITE | N/A | N/A | N/A | 178.3297, 170.39, use C-G, not for articles used for holding food during cooking |

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.