

# Certificate of Analysis

## Product Information:

<b>Part Number</b>	910001
<b>Description</b>	PH ELECTRODE STORAGE SOLUTION, 475ML
<b>Lot Number</b>	SST5
<b>Expiration Date</b>	NA; RECOMMENDATION USAGE 2 YEARS FROM DATE OF RECEIPT
<b>Method of Analysis</b>	POTENTIOMETRIC

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
3.0 M +/- 0.1M POTASSIUM CHLORIDE	2.9	M POTASSIUM CHLORIDE
6.75+/-0.20 PH UNITS AT 25 ° C	6.68	pH @ 25° C

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM#</u>
Potassium Dihydrogen Phosphate	186-I-g
Disodium Hydrogen Phosphate	186-II-g
Sodium Bicarbonate	191d-I
Sodium Carbonate	191d-II
Sodium Chloride	2201

**Certificate Date:**

31-Jul-14

**Approved By:**



**Quality Representative**

[thermoscientific.com/water](http://thermoscientific.com/water)

© 2012 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific Inc. & its subsidiaries

Water Analysis Instruments

**North America**  
Toll Free: 1-800-225-1480  
Tel: 1-978-232-6000  
[info.water@thermo.com](mailto:info.water@thermo.com)

**Netherlands**  
Tel: (31) 033-2463887  
[info.water.uk@thermo.com](mailto:info.water.uk@thermo.com)

**China**  
Tel: (86) 21-68654588  
[wai.asia@thermofisher.com](mailto:wai.asia@thermofisher.com)

**India**  
Tel: (91) 22-4157-8800  
[wai.asia@thermofisher.com](mailto:wai.asia@thermofisher.com)

**Singapore**  
Tel: (65) 6778-6876  
[wai.asia@thermofisher.com](mailto:wai.asia@thermofisher.com)

**Japan**  
Tel: (81) 045-453-9175  
[wai.asia@thermofisher.com](mailto:wai.asia@thermofisher.com)

**Australia**  
Tel: (613) 9757-4300  
in Australia (1300) 735-295  
[InfoWaterAU@thermofisher.com](mailto:InfoWaterAU@thermofisher.com)

**Thermo**  
SCIENTIFIC  
Part of Thermo Fisher Scientific