

DECLARATION OF CONFORMITY

Thermo Environmental Instruments, Inc. located at 27 Forge Parkway, Franklin, MA 02038, USA, declares under our sole responsibility that the following products:

Thermo Scientific models ADR 1500 Particulate Monitor

To which this declaration relates are in conformity with the following standards or other normative documents:

UL 61010-1

Issued: 2012/05/11 Ed: 3 UL Standard for Safety Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements

CAN/CSA C22.2 #61010-1

Issued: 2012/05/11 Ed: 3 Standard for Safety Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements

IEC/EN 61010-1

Issued: 2010 (3rd Edition) Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use Part 1: General Requirements

FCC 47CFR 15B cIB

Issued: 2007/10/01 Title 47 CFR Part 15 Subpart B Unintentional Radiator Class B Verification

CENELEC EN 61326-1

Issued: 2006 Electrical equipment for measurement, control and laboratory use EMC requirements Part 1: General requirements

I, the undersigned, hereby declare that the equipment specified above conforms to the above Directive(s) and Standard(s).

ShiMin Chen
Manager, Electrical Engineering
Thermo Fisher Scientific, Inc.
27 Forge Parkway
Franklin, MA 02038 USA
(508) 520-0430