

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 158**

of **Life Technologies Europe B.V.**
European Calibration Services (ECS)

This annex is valid from: **15-05-2024** to **01-07-2027**

Replaces annex dated: **10-04-2024**

Location(s) where activities are performed under accreditation

Head Office

Kwartsweg 2
 2665 NN
 Bleiswijk
 The Netherlands

Location	Abbreviation/ location code
Kwartsweg 38 2665 NN Bleiswijk The Netherlands	BL

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
LF 0 0	DC/LF Quantities				
LF 1 1	DC Voltage			generating	BL
	3.300 V		0.002 V		
	33.00 V		0.02 V		
	330.0 V		0.1 V		
	1 kV		1 V		
	DC Millivolt			generating	BL
	33.0 mV		0.1 mV		
	330.0 mV		0.2 mV		

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

of **Life Technologies Europe B.V.**
European Calibration Services (ECS)

This annex is valid from: **15-05-2024** to **01-07-2027**

Replaces annex dated: **10-04-2024**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
LF 2 1	DC MilliAmp			generating	BL
	33.00 mA		0.03 mA		
	330.0 mA		0.7 mA		
	DC MicroAmp			generating	BL
	330.0 µA		0.3 µA		
	3300 µA		3 µA		
	DC Amp			generating	BL
	3.000 A		0.005 A		
	10.00 A		0.02 A		
LF 3 1	AC Voltage			generating	BL
	330.0 mV	60 Hz	0.7 mV		
	600.0 mV	13 kHz	1.6 mV		
	3.300 V	60 Hz	0.007 V		
	3.300 V	20 kHz	0.015 V		
	33.00 V	60 Hz	0.04 V		
	33.00 V	20 kHz	0.10 V		
	330.0 V	60 Hz	0.5V		
	330.0 V	2.5 kHz	0.7 V		
	500 V	60 Hz	1 V		
	1 kV	1 kHz	2V		
LF 4 1	AC MilliAmp			generating	BL
	33.00 mA	60 Hz	0.12 mA		
	330.0 mA	60 Hz	1.6 mA		

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 158**

of **Life Technologies Europe B.V.**
European Calibration Services (ECS)

This annex is valid from: **15-05-2024** to **01-07-2027**

Replaces annex dated: **10-04-2024**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
LF 4 1	AC MicroAmp			generating	BL
	330.0 μ A	60 Hz	1.1 μ A		
	3300 μ A	60 Hz	11 μ A		
LF 4 1	AC Amp			generating	BL
	3.000 A	60 Hz	0.008 A		
LF 6 1	Resistance			generating	BL
	330.0 Ω		0.5 Ω		
	3.300 k Ω		0.003 k Ω		
	33.00 k Ω		0.03 k Ω		
	330.0 k Ω		0.3 k Ω		
	3.300 M Ω		0.005 M Ω		
	30.00 M Ω		0.07 M Ω		
TF 0 0	Time and Frequency				
TF 2 1	Frequency			generating	BL
	60.00 Hz		0.01 Hz	U _{nom} = 110 V	
	50.00 Hz		0.01 Hz	U _{nom} = 230 V	

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 158**

of **Life Technologies Europe B.V.**
European Calibration Services (ECS)

This annex is valid from: **15-05-2024** to **01-07-2027**

Replaces annex dated: **10-04-2024**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ²	Remarks	Location
TE 0 0	Temperature				
TE 4 1	Thermometers	4.00 °C – 96.00 °C	0.09 °C	In water bath with multi-well dry block inserts	BL
		30.00 °C – 115.00 °C	0.10 °C	In silicon oil bath with multi-well dry block inserts	

Remarks:

Environmental conditions $T = (23 \pm 3) \text{ °C}$, $RH = (50 \pm 25) \text{ \% rh}$ for Electrical
 $T = (23 \pm 2) \text{ °C}$, $RH = (50 \pm 25) \text{ \% rh}$ for Thermometers (for water bath)
 $T = (23 \pm 5) \text{ °C}$, $RH = (50 \pm 25) \text{ \% rh}$ for Thermometers (for silicon bath)

Calibrations are performed inside the laboratory, unless specified otherwise.

² Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, U , is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".