



CERTIFICATE OF ANALYSIS

PRODUCT MT0249
BHI SLANT W/5% SHEEP BLOOD 16X125MM

LOT NUMBER 0305

EXPIRY DATE 27.12.2010
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Cherry red slant	Conforms to Specification
pH	7.50	7.20 - 7.60
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C± 2°C for 5days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Trichophyton mentagrophytes ATCC® 9533	Growth	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Growth	Conforms to Specification
Neisseria meningitidis ATCC® 13090	Growth	Conforms to Specification
Streptococcus pneumoniae ATCC® 6305	Growth	Conforms to Specification
Candida albicans ATCC® 60193	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager