



Certificate of Quality

Product Name: M5(3ml) PK/72
Lot Number: 055984

Product Number: R12515
Expiration Date: 2012-12-06
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Maintenance:		
Chlamydia trachomatis	VR-880	Pass
Mycoplasma hominis	23114	Pass
Mycoplasma pneumoniae	15331	Pass
Ureaplasma urealyticum	27618	Pass
Herpes simplex Type 2	VR-734	Pass
Cytomegalovirus	VR-977	Pass
Respiratory syncytial Virus	VR-1302	Pass

Toxicity:		
MRC5 Cells	CCL-171	Pass
VERO Cells	CCL-81	Pass
McCoy Cells	RLL-169	Pass

Physical Characteristics:	
Appearance:	Orange-Pink Clear Broth w/ Clear Beads
pH @ 20-25°C:	7.1 - 7.5

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor