

# CERTIFICATE OF ANALYSIS

**PRODUCT** TM1300S5  
SIMMONS CITRATE AGAR (5ML)

**LOT NUMBER** 1004902

**EXPIRY DATE** 08.06.2011

---

Physical Characteristics	Results	Specification
Appearance	Dark green coloured gel	Dark green coloured gel
pH	6.87	6.80 - 7.20

---

**Sterility:** Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

---

Microbiological Characteristics	Results	Specification	Quantitative Tests
<b>O2, 35°C, 24hr</b>			
<i>Enterobacter aerogenes</i> ATCC®13048	Positive, medium changed to blue color		
<i>Escherichia coli</i> ATCC®25922	Negative, medium remains green		

---

**Released By**



**Gan May Ping**

---

**MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)**

**(Formerly known as Biomedica Laboratories Sdn Bhd)**

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: [oxid.myqcbml@thermofisher.com](mailto:oxid.myqcbml@thermofisher.com)