

Tel 585-586-8800 Fax 585-899-7605 75 Panorama Creek Drive, Rochester, NY 14625 Product Certificate
Thermo Scientific
Nalgene and Nunc Products

Thermo Fisher Scientific hereby certifies that the product identified below is produced, inspected and found to be in compliance with product and quality specification requirements as documented in our ISO 13485:2003 Quality Management System (QMI-SAI Global File No. 1606319 and 1606321) in the USA.

RIDM Reund Robert Prescott Mgr. QA/RA

The following information represents Product Certification for: Item#: 2148-05

Certificate issued: 02/25/2011

Description: RB,PETG,2.5X,QUICK ACTN CLOS Lot#: 1040436 Use Before: 02/28/2016 Manufactured: 02/14/2011

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-0449-91	CLOS,48MM,Q/A,W/O SLRNG,PP,WHT	COMPONENT PART				
8-0028-16P	RESIN,PPCO,RAD STAB,WHITE,INJ	COLOR MIX (PPCO, RAD STAB,WHT)	N/A	PASSED	PASSED	N/A
8-0028-04	RESIN,PPCO,RAD STER,INJ	POLYPROPYLENE COPOLYMER	7478	PASSED	PASSED	177.1520 (a)(3)(i) & (c)3.1(a)except for cooking, (useconditions C-H)
8-0099-34	COLOR,WHT,MULTI	COLORANT, WHITE	16513	PASSED	PASSED	177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210
1-0449-88P	BTL,ROLL,2.5X,Q/AL,PETG,SM RIB	COMPONENT PART				
8-0001-32	RESIN, PETG, IBM, EBM, INJ	COPOLYESTER, PETG NATURAL	9987	PASSED	PASSED	177.1315(b)(1) and 174.5

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Product was Gamma Irradiation Sterilized. Product was dosimetric released per ANSI/AAMI/ISO 11137 guidelines. Product was determined to be non-pyrogenic at a level < 0.5 EU/ml as by "Guidelines on the Validation of the Limulus Amebocyte Lysate Test," as defined by the FDA (12/87), as an end product endotoxin test for human and animal parenteral drugs, biological products and medical devices.

The foam liner and PP closure materials have been tested and shown to comply with USP Class VI requirements and also have been shown to be non-cytotoxic. The liner material meets requirements of CFR 21, Sections 175.300, 177.1520, 177.1210, 178.2010 of the Federal Food and Drug Act. The closure materials meet requirements of CFR 21, Section 177.1520 of the Federal Food and Drug Act.