



## CERTIFICATE OF ANALYSIS

**PRODUCT** MT0326  
CHOCOLATE AGAR SLANT 5ML.16x100MM

**LOT NUMBER** 1041

**EXPIRY DATE** 02.06.2011  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Chocolate Slant	Conforms to Specification
pH	7.30	7.00 - 7.40
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Neisseria meningitidis ATCC® 13090	Growth	Conforms to Specification
Neisseria gonorrhoeae ATCC 43069	Growth	Conforms to Specification
Haemophilus influenzae ATCC® 10211	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager