

Tel 585-586-8800 Fax 585-899-7605 75 Panorama Creek Drive, Rochester, NY 14625 Product Certificate
Thermo Scientific
Nalgene and Nunc Products

Thermo Fisher Scientific hereby certifies that the product identified below is produced, inspected and found to be in compliance with product and quality specification requirements as documented in our Quality Management System in Monterrey, Mexico.

RADA Reuse Robert Prescott Mgr. QA/RA

The following information represents Product Certification for: Item#: 3662-0250

Certificate issued: 08/24/2011

Description: GRADUATED CYLINDR PP;250ML Lot#: 1054025

Manufactured: **08/16/2011**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-3205-45	BASE,250ML,GRAD,CYL,PP,BLUE	COMPONENT PART				
8-0029-14P	RESIN,PPCO,BLU,INJ	COLOR MIX (PPCO, BLU, RESIN)	N/A	N/A	N/A	N/A
8-0029-04	RESIN, PP, COPOLYMER, INJ	POLYPROPYLENE COPOLYMER	1698	N/A	N/A	177.1520(a)(3)(i) and(c)3.1a, 3.2a
8-0099-74	COLOR, BLUE, MULTI	COLORANT, BLUE, MULTI-RESIN	N/A	PASSED	PASSED	177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210
1-3205-65P	BODY,MOLDED,250ML,GRAD,CYL, PP	COMPONENT PART				
8-0028-05	RESIN, PP, COPOLYMER, INJ	POLYPROPYLENE COPOLYMER	7478	N/A	PASSED	177.1520(a)(3)(i), (c)3.1a,except for cooking, (useconditions C-H)

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

NALGENE Graduated Cylinders are designed to meet accuracy requirements of ASTM Class B, E1272, "Cylinder, Graduated, Laboratory, Glass" and all requirements of ISO Standard 6706, "Plastic Laboratory Ware- Graduated Measuring Cylinders."