



CERTIFICATE OF ANALYSIS

PRODUCT **BO0556T**
YEAST EXTRACT AGAR
10X250ML IN 300ML SIROP

LOT NUMBER 1055270

EXPIRY DATE 2009.07.04

Delivery/Customer information

Date Printed
2013.11.11
Delivery No.

Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Straw 2
pH (25°C)	7.1	7.0 - 7.4
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Escherichia coli ATCC@25922	43	conforms	Pale straw cols
Pseudomonas aeruginosa ATCC@27853	36	conforms	Straw cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd