

**CERTIFICATE OF ANALYSIS**

**PRODUCT** PP2311  
MRS AGAR PLATE

**LOT NUMBER** 1058270

**EXPIRY DATE** 2008.09.23

**RELEASE DATE** 2008.07.31

---

**Physical Testing** **Results** **Specification***Tested in accordance with Oxoid Method QC13 unless otherwise stated*

Appearance (Colour)	Pass	Beige
Clarity	Pass	Translucent
Gel Strength	Pass	Satisfactory
pH	6.0	6.0 - 6.4
Sterility @ 30°C for 3 days /sample plan	Pass	Within Limits

*This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1*

---

**Microbiology Testing** **Result** **Specification** **Method**

Lactobacillus fermentans ATCC® 9338	Pass	Beige colonies	
Escherichia coli ATCC® 25922	Pass	Inhibited	QC11
Staphylococcus aureus ATCC® 25923	Pass	Inhibited	QC11

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Rina Pattullo



---

**Approved Signatory**

Accredited for compliance with ISO/IEC 17025.  
This report may not be reproduced except in full.



Accreditation  
Number 3186



Accreditation No. 3186

**MANUFACTURED AND TESTED BY:****Thermo Fisher Scientific, Microbiology Division**

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054

Email: [micro\\_orders@thermofisher.com](mailto:micro_orders@thermofisher.com) Website: [www.thermofisher.com.au](http://www.thermofisher.com.au)