



CERTIFICATE OF ANALYSIS

PRODUCT TM1564
LACTRITMEL+ANTI+ACTI SLOPE 12ML

LOT NUMBER 1063561

EXPIRY DATE 2009.09.16

RELEASE DATE 2008.09.24

Physical Testing **Results** **Specification**
Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	White
Gel Strength	Pass	Satisfactory
Clarity	Pass	Translucent
pH	6.3	6.1 - 6.5

Microbiological Testing	Result	Specification	Method
Escherichia coli ATCC® 25922	Pass	No growth	
Staphylococcus aureus ATCC® 25923	Pass	No growth	
Proteus mirabilis ATCC® 12453	Pass	No growth	
Candida albicans ATCC® 10231	Pass	White colonies	
Trichophyton mentagrophytes MVQC 0104	Pass	White fluffy	
Trichophyton rubrum MVQC 0103	Pass	White fluffy, red reverse pigment	QC03
Aspergillus brasiliensis ATCC® 16404	Pass	No growth	

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Rina Pattullo

.....
Approved Signatory

Accredited for compliance with ISO/IEC 17025.
 This report may not be reproduced except in full.



MANUFACTURED AND TESTED BY: Oxoid Australia Pty Limited

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054

Email: oxoid.au@thermofisher.com Website: www.oxoid.com.au