

Thermo Fisher Scientific hereby certifies that the product identified below is produced, inspected and found to be in compliance with product and quality specification requirements as documented in our Quality Management System in Monterrey, Mexico.



Lisa Adams  
Mgr. Quality Engineering

The following information represents Product Certification for: Item#: **330821**

Certificate issued: **04/14/2012**

Description: **MICRO MAX CRYOSTORE BOX PC**

Lot#: **1069813**

Manufactured: **04/04/2012**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
009274A00P	CRYOBOX LID, POLYCARBONATE	COMPONENT PART				
8-0056-46P	RESIN,PC,OFF-WHT,INJ	COMPONENT PART				
8-0056-53	RESIN,PC,INJ	POLYCARBONATE, INJ./IBM	N/A	PASSED	PASSED	177.1580
8-0099-99	COLOR,OFF-WHITE,PC,14145MR	COLOR, OFF-WHITE, PC	N/A	N/A	N/A	Not FDA
009274B00P	CRYOBOX BASE,POLYCARBONATE	COMPONENT PART				
8-0056-46P	RESIN,PC,OFF-WHT,INJ	COMPONENT PART				
8-0056-53	RESIN,PC,INJ	POLYCARBONATE, INJ./IBM	N/A	PASSED	PASSED	177.1580
8-0099-99	COLOR,OFF-WHITE,PC,14145MR	COLOR, OFF-WHITE, PC	N/A	N/A	N/A	Not FDA

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.