



## CERTIFICATE OF ANALYSIS

**PRODUCT** MT1660  
**PHYTONE YEAST EXTRACT AGAR PCK 25**  
**15 ML 28 X 95MM PCK 25**

**LOT NUMBER** 1075

**EXPIRY DATE** 15.09.2011  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear amber	Conforms to Specification
pH	6.70	6.40 - 6.80
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
Candida albicans ATCC® 10231	Growth	Conforms to Specification
Trichophyton mentagrophytes ATCC® 9533	Growth	Conforms to Specification
A. brasiliensis frm. niger ATCC® 16404	Growth	Conforms to Specification
Escherichia coli ATCC® 25922	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition	Conforms to Specification
Trichophyton verrucosum ATCC® 42898	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager