

**CERTIFICATE OF ANALYSIS**

**PRODUCT** SR0244E  
**OCLA (ISO)DIFFERENTIAL SUPPLEMENT**

**LOT NUMBER** 1077476

**EXPIRY DATE** 2012.05.31

**DATE OF MANUFACTURE** 2011.08.02

## Delivery/Customer information

Date Printed  
 2014.01.03  
 Delivery No.

Customer  
 Customer Order number

Physical Characteristics	Results	Specification
Appearance	Opaque, cream liquid	Opaque, cream liquid

  

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	----------------	-------------	--------------------	-------------

**Tested in Chromogenic Listeria Agar Base (ISO) CM1084 and OCLA (ISO) Selective Supplement SR0226****Aerobic incubation at 37°C for 48 hours**

<i>Listeria monocytogenes</i> ATCC@13932	62	55	89	Blue cols with halo
<i>Listeria innocua</i> ATCC@33090	61	62	102	Blue cols w/o halo

Control Medium: Columbia Blood Agar Base enriched with 5% v/v horse blood

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu).

Refer to product specification for full details.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. The results reported were obtained at the time of release.

Lot Accepted. 2011.08.08

Carol MacBride  
 PRODUCT PERFORMANCE MANAGER

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

**OXOID LIMITED**

Wade Road, Basingstoke, Hampshire RG24 8PW. England  
 oxid@thermofisher.com www.oxid.com