



Quality Assurance Certificate

Code: BC0100M
Product: SIGNAL BLOOD CULTURE SYSTEM 1 X 20
Lot: 1077581
Expiry: 31 August 2013
Components:
BC0101M: 20110409P

PHYSICAL CHARACTERISTICS: As appropriate tests on moisture, appearance, pH, clarity and gel strength were satisfactory.

BIOLOGICAL PERFORMANCE: Satisfactory performance of:

| | |
|----------------------------|------------|
| Clostridium perfringens | ATCC 13124 |
| Bacillus cereus | ATCC 10876 |
| Bacteroides fragilis | ATCC 25285 |
| Clostridium novyi | ATCC 27606 |
| Escherichia coli | ATCC 25922 |
| Fusobacterium nucleatum | ATCC 10953 |
| Haemophilus influenzae | ATCC 19418 |
| Klebsiella pneumoniae | ATCC 29665 |
| Neisseria meningitidis | ATCC 13077 |
| Pseudomonas aeruginosa | ATCC 27853 |
| Staphylococcus aureus | ATCC 25923 |
| Staphylococcus epidermidis | ATCC 14990 |
| Streptococcus pneumoniae | ATCC 6303 |
| Streptococcus mutans | ATCC 25175 |
| Candida albicans | ATCC 10231 |
| Prevotella bivia | ATCC 29303 |

Date: 11 November 2013

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

The information given above is believed to be correct, however, both the information and product are offered without warranty for any specific use, other than that specified in the current Oxoid Manual.

C. MACBRIDE
Product Performance Manager

C. BOOTH
V.P. Science and Technology



Certificate Nos. FM09914
BS EN ISO 9001
BS EN ISO 13485

OXOID LIMITED
Wade Road, Basingstoke, Hants RG24 8PW, England
Oxoid@Oxoid.com www.oxoid.com

FDA Reg. No. 8010096