



CERTIFICATE OF ANALYSIS

PRODUCT MT0514
VRE SELECTIVE BROTH(BEA) 10 ML 100 PCK
16 X 125 MM CR00575

LOT NUMBER 1081

EXPIRY DATE 21.09.2011
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

| Physical Characteristics | Results | Specification |
|------------------------------------|-------------|---------------------------|
| Appearance | Clear amber | Conforms to Specification |
| pH | 7.10 | 6.90 - 7.30 |
| Sterility @ 35 °C ± 2°C for 3 days | Conforms | Conforms to specification |

| Microbiological Performance Organism Tested | Expected Results | Specification |
|--|------------------------------|---------------------------|
| Pseudomonas aeruginosa ATCC® 27853 | Inhibition | Conforms to Specification |
| Escherichia coli ATCC® 25922 | Inhibition | Conforms to Specification |
| Enterococcus faecalis ATCC® 29212 | Inhibition | Conforms to Specification |
| Enterococcus faecalis ATCC® 51299 | Growth (blackening of broth) | Conforms to Specification |

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager