

**CERTIFICATE OF ANALYSIS**

**PRODUCT** PP2369  
HBA/BRILLIANCE UTI CLARITY

**LOT NUMBER** 1086195

**EXPIRY DATE** 2009.05.29

**RELEASE DATE** 2009.04.20

**Physical Testing Side 1** **Results** **Specification**  
*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

Appearance (Colour) - HBA	Pass	Red
Clarity - HBA	Pass	Opaque
Gel Strength	Pass	Satisfactory
pH - HBA	7.4	7.2 - 7.6

**Microbiological Testing Side 1** **Result** **Specification** **Method**

Escherichia coli ATCC® 25922	Pass	Grey, wet colonies	
Staphylococcus aureus ATCC® 25923	Pass	white/cream colonies	
Streptococcus pyogenes ATCC® 19615	Pass	Beta haemolytic colonies	

**Physical Testing Side 2** **Results** **Specification**  
*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

Appearance (Colour) - Chrom	Pass	Colourless/Straw
Clarity - Chrom	Pass	Translucent
Gel Strength	Pass	Passed
pH - Chrom	7.1	6.8 - 7.2

**Microbiological Testing Side 2** **Result** **Specification** **Method**

Escherichia coli ATCC® 25922	Pass	Pink colonies	
Staphylococcus aureus ATCC® 25923	Pass	Cream colonies	
Klebsiella pneumoniae ATCC® 13883	Pass	Blue colonies	
Proteus mirabilis ATCC® 12453	Pass	Beige colonies, no swarming	
Enterococcus faecalis ATCC® 29212	Pass	Blue/green colonies	
Staphylococcus saprophyticus ATCC® 1530	Pass	Pink/white colonies	



.....  
**Approved Signatory**

Accredited for compliance with ISO/IEC 17025.  
This report may not be reproduced except in full.



Number 3186



Accreditation No. 3186

**MANUFACTURED AND TESTED BY:**

**Thermo Fisher Scientific, Microbiology Division**

20 Dalgleish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054

Email: [micro\\_orders@thermofisher.com](mailto:micro_orders@thermofisher.com) Website: [www.thermofisher.com.au](http://www.thermofisher.com.au)

## CERTIFICATE OF ANALYSIS

---

<b>Sterility Testing</b>	<b>Results</b>	<b>Specification</b>
--------------------------	----------------	----------------------

---

*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

---

Sterility @ 30°C for 3 days /sample plan	Pass	Within Limits
--	------	---------------

*This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1*

---

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Rina Pattullo

**MANUFACTURED AND TESTED BY:**

**Thermo Fisher Scientific, Microbiology Division**

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054

Email: [micro\\_orders@thermofisher.com](mailto:micro_orders@thermofisher.com) Website: [www.thermofisher.com.au](http://www.thermofisher.com.au)