



CERTIFICATE OF ANALYSIS

PRODUCT **BO0350S**
 HALF FRASER BROTH
 10X225ML IN 250ML PET

LOT NUMBER 1116816

EXPIRY DATE 2010.07.20

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	conforms	Straw 4 with blue iridescence
pH (25°C) 7.2		7.0 - 7.4
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test Result	Specification
Listeria monocytogenes ATCC@13932	< 1000	conforms	Aesculin hydrolysis (blackening)
Listeria monocytogenes ATCC35152	< 1000	conforms	Aesculin hydrolysis (blackening)
Enterococcus faecalis ATCC@29212	33	conforms	No aesculin hydrolysis (no blackening)
Bacillus cereus ATCC@10876	< 1000	conforms	No aesculin hydrolysis (no blackening)
Escherichia coli ATCC@25922	< 100000	conforms	No aesculin hydrolysis (no blackening)

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

M. E. McCallum

MOIRA MCCALLUM
 PRODUCT PERFORMANCE MANAGER, G&M
 PROCTER LTD



G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,
 4 Auld Bond Road, Perth., PH1 3FX