

Tel 585-586-8800 Fax 585-899-7605 75 Panorama Creek Drive, Rochester, NY 14625



Thermo Fisher Scientific hereby certifies that the product identified below is produced, inspected and found to be in compliance with product and quality specification requirements as documented in our ISO 13485:2003 Quality Management System (QMI-SAI Global File No. 1616978) in Monterrey, Mexico.

Tresia O'Shea Manager, Regulatory Compliance

The following information represents Product Certification for: Item#: 155-0045

Certificate issued: 03/21/2014

Description: FLT,SFCA,45MM150ML,50MM;.45U= =

Lot#: 1118207

Use Before: 03/28/2019

Manufactured: 03/12/2014

| Part Number | Description | Common Name | DMF# | Cytotoxicity | USP Class VI | FDA Compliance - 21 CFR |
|-------------|-------------------------------------|------------------------------|-------|--------------|--------------|--|
| 1-0014-83 | COLLAR, FLT, 45/50, MF75, HIPS, YEL | COMPONENT PART | | | | |
| 8-0077-11P | RESIN, PS, HIGH IMPACT, YEL, INJ | HIGH IMPACT PS, YELLOW | N/A | PASSED | PASSED | N/A |
| 8-0077-13 | RESIN, PS, HIGH IMPACT, INJ | POLYSTYRENE | 1623 | PASSED | PASSED | 177.1640 |
| 8-0099-32 | COLOR, YEL | COLORANT, YELLOW | N/A | PASSED | PASSED | 177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210 use conditions B-H |
| 1-0014-95 | ADP,FLT,Q/D,LDPE,WHT,MF-75 | COMPONENT PART | | | | |
| 8-0049-14P | RESIN, LDPE, WHT, INJ | COLOR MIX (LDPE, WHT, INJ) | N/A | N/A | N/A | N/A |
| 8-0049-01 | RESIN, LDPE, INJ | LOW-DENSITY POLYETHYLENE | 9522 | PASSED | PASSED | 177.1520(a)(2)(i) |
| 8-0099-26 | COLOR, WHT, OLF | COLORANT, WHITE | N/A | N/A | N/A | 178.3297, 170.39, use C-G, notfor articles used for holdingfood during cooking |
| 1-1803-52 | CLOS,45/430,HDPE,WHT,NALGE | COMPONENT PART | | | | |
| 8-0042-16P | RESIN, HDPE, WHT, INJ | COLOR MIX (RESIN, HDPE, WHT) | N/A | PASSED | PASSED | N/A |
| 8-0042-01 | RESIN, HDPE, INJ | HIGH-DENSITY POLYETHYLENE | 1646 | PASSED | PASSED | 176.170(c), 177.1520(c)3.2a |
| 8-0099-34 | COLOR, WHT, MULTI | COLORANT, WHITE | 16513 | PASSED | PASSED | 177.1350, 1520, 1620,178.3297, 181.28 |
| 1-0014-62 | LOWER, FLT, 150ML, 45MM NK, ANON | COMPONENT PART | | | | |
| 8-0077-12 | RESIN, PS, INJ, IBM | POLYSTYRENE | 1623 | PASSED | PASSED | 177.1640 |
| 1-0014-42 | COV,FLT,150/250mL,THERMO | COMPONENT PART | | | | |
| 8-0077-12 | RESIN, PS, INJ, IBM | POLYSTYRENE | 1623 | PASSED | PASSED | 177.1640 |
| 1-0014-92 | UPPER, FLT, 150 ML, MF-75, PS | COMPONENT PART | | | | |
| 8-0077-12 | RESIN, PS, INJ, IBM | POLYSTYRENE | 1623 | PASSED | PASSED | 177.1640 |



Tel 585-586-8800 Fax 585-899-7605 75 Panorama Creek Drive, Rochester, NY 14625



| Thermo Fisher Scientific hereby certifies that the product identified below is produced, inspected and found to |
|---|
| be in compliance with product and quality specification requirements as documented in our ISO 13485:2003 |
| Quality Management System (QMI-SAI Global File No. 1616978) in Monterrey, Mexico. |

Tresia O'Shea Manager, Regulatory Compliance

The following information represents Product Certification for: Item#: 155-0045

Certificate issued: 03/21/2014

Description: FLT,SFCA,45MM150ML,50MM;.45U= = Lot#: 1118207 Use Before: 03/28/2019 Manufactured: 03/12/2014

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Product was Gamma Irradiation Sterilized. Product was dosimetric released per ANSI/AAMI/ISO 11137 guidelines. Product was determined to be non-pyrogenic at a level < 0.25 EU/ml per USP < 85 > .