



CERTIFICATE OF ANALYSIS

PRODUCT TM0118
NITRATE BROTH 3ml

LOT NUMBER 1120938

EXPIRY DATE 2011.02.22

RELEASE DATE 2010.02.23

Physical Testing **Results** **Specification**
Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	Colourless / pale yellow
Clarity	Pass	Translucent
pH	7.4	7.2 - 7.6

Microbiological Testing	Result	Specification	Method
Escherichia coli ATCC® 25922	Pass	Positive, red	QC27
Alcaligenes faecalis ATCC® 8750	Pass	Colourless -ve / Zinc red -ve	QC27

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

The product is stable to the stated expiry date when stored in accordance with instructions per product label.

Released by: Paul Costello

.....
Approved Signatory

Accredited for compliance with ISO/IEC 17025.
This report may not be reproduced except in full.



MANUFACTURED AND TESTED BY: Oxoid Australia Pty Limited
 20 Dalglish Street, Thebarton, South Australia 5031
 Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054
 Email: oxoid.au@thermofisher.com Website: www.oxoid.com.au