



## CERTIFICATE OF ANALYSIS

Page 1 of 1

**PRODUCT** TM0080  
CSF BROTH 12ML

**LOT NUMBER** 1132554

**EXPIRY DATE** 2011.06.01

**RELEASE DATE** 2010.06.07

---

Physical Testing	Results	Specification
------------------	---------	---------------

*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

Appearance (Colour)	Pass	Brown
Clarity	Pass	Translucent
pH	7.4	7.0 - 7.4
Sterility @ 30°C for 7days	Pass	Within Limits

*This test is performed in accordance Oxoid method QC21, NATA FADs, ASM guidelines: Normal inspection AQL 2.5, General inspection level=1*

---

Microbiological Testing	Result	Specification	Method
Haemophilus influenzae ATCC® 10211	Pass	Growth	QC16
Streptococcus pneumoniae ATCC® 6305	Pass	Growth	QC16
Streptococcus pyogenes ATCC® 19615	Pass	Growth	QC16
Neisseria meningitidis ATCC® 13090	Pass	Growth	QC16

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

The product is stable to the stated expiry date when stored in accordance with instructions per product label.

Released by: Paul Costello

*Paul Costello*

.....  
**Approved Signatory**

Accredited for compliance with ISO/IEC 17025.  
This report may not be reproduced except in full.



**MANUFACTURED AND TESTED BY: Oxoid Australia Pty Limited**

20 Dalgleish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054

Email: [oxoid.au@thermofisher.com](mailto:oxoid.au@thermofisher.com) Website: [www.oxoid.com.au](http://www.oxoid.com.au)