



Certificate of Analysis

CloneMiner™ II cDNA Library Construction Kit

Product No. A11180-01

Lot No. 1138379

Titer Determination:

Specification: Total cfu from Test cDNA BP reaction must be $\geq 1.0 \times 10^7$ cfu.

Result: Meets specifications.

PCR Full Length Analysis:

Specification: $\geq 87\%$ of 24 colonies must have an insert > 500 bp.

Result: Meets specifications.

All products are for research use only. Caution: Not intended for human or animal diagnostic or therapeutic uses. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com

A handwritten signature in black ink, appearing to read "Rick Fletcher".

Rick Fletcher
Director, Quality
Issued on 19 Oct 2012