



CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT PO5176A
BRILLIANCE (TM) E.COLI/COLIFORM SELECTIV

LOT NUMBER 1146358

EXPIRY DATE 2012.03.16

TYPICAL FORMULATION

	grams per litre
Peptone	8.00000
Disodium hydrogen phosphate	2.20000
Sodium chloride	5.00000
Potassium dihydrogen phosphate	1.80000
Sodium lauryl sulphate	0.10000
Chromogenic Mix	0.35000
Agar	10.60000
Tryptophane	1.00000

General Characteristics	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Transparent
pH	6.9	6.5 - 6.9
Packaging/Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Escherichia coli ATCC®25922	82	47	1mm violett colonies
Citrobacter freundii ATCC®8090	<10000	Conforms	Good growth, small rose colonies
Staphylococcus aureus ATCC®25923	6.2E+04	0	Partial inhibition (<= 100 cfu)

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature