

Certificate of Analysis

Product Name: Streptex Extraction Enzyme KT/25TST
Product Code: R30951001
Lot Number: 1146691
Expiration Date: 31.03.2015
Date of Manufacture: 15.05.2012

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W73UZL55T	ZL55 Extraction Enzyme Freeze dried	1084904	31.03.2015

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Efficacy of the Extraction Enzyme	Each Latex must show obvious agglutination with its homologous extract in the 10 and 60 minute extraction procedures	Conforms
	Each Latex must not show agglutination with the heterologous extracts although a trace granularity is acceptable	Conforms
Latex Reactivity with Extraction Enzyme	The Extraction Enzyme must show no agglutination with the Latexes , although a trace reaction is acceptable	Conforms



Fay Curd
 Quality Assurance Manager