

# CERTIFICATE OF ANALYSIS

**PRODUCT** TM1020S5  
BILE ESCULIN AGAR (5ML)

**LOT NUMBER** 1151018

**EXPIRY DATE** 19.04.2012

---

Physical Characteristics	Results	Specification
Appearance	Straw brown coloured gel	Straw brown coloured gel
pH	7.07	6.90 - 7.30

---

**Sterility:** Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

---

Microbiological Characteristics	Results	Specification	Quantitative Tests
<b>O2, 35°C, 24-48hr</b>			
<i>Enterococcus faecalis</i> ATCC®29212	Black coloured medium		
<i>Streptococcus pyogenes</i> ATCC®19615	Complete inhibition		

---

**Released By**



**Zulhusni Amir**

---

**MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)**

**(Formerly known as Biomedica Laboratories Sdn Bhd)**

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: [oxid.myqcbml@thermofisher.com](mailto:oxid.myqcbml@thermofisher.com)