

Product : PBS-Trion
Catalogue Number : 04194807
Lot Number : 1185268
Expiry Date : 07/2014
Country of Origin : UK

TEST	SPECIFICATION	RESULT	UNITS
a01 Bacterial	Negative	Negative	
a02 Fungal	Negative	Negative	
a96 Endotoxin	0.001 to 0.250	0.040	eu/mL
b01 pH	7.00 to 7.30	7.11	pH Units
b02 Osmolality	270 to 310	281	mOs/kg H2O
b33 Volume	Satisfactory	Satisfactory	
d07 Maintenance	Satisfactory	Satisfactory	
q01 Documentation	Satisfactory	Satisfactory	
q02 Appearance (Liquid)	Satisfactory	Satisfactory	

End of Report

Life Technologies Limited
 3 Fountain Drive
 Inchinnan Business Park
 Paisley PA4 9RF UNITED KINGDOM

Q. A. Approved

By *Laura McEllen*

Date: 19-Jul-2012