



CERTIFICATE OF ANALYSIS

PRODUCT TM0543
POTASSIUM TELLURITE 1% 100ML

LOT NUMBER 1200672

EXPIRY DATE 2012.07.04

RELEASE DATE 2012.01.04

Physical Testing **Results** **Specification**

Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	Colourless
Clarity	Pass	Translucent
Sterility @ 30°C for 2 days	Pass	Within Limits

This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1

Microbiological Testing **Result** **Specification** **Method**

Coryn. diphtheriae gravis ATCC® 19409	Pass	Black colonies	QC27
Coryn. diphtheriae intermedius ATCC®1477	Pass	Black colonies	QC27
Staphylococcus epidermidis ATCC® 12228	Pass	Inhibited	QC27

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Rina Pattullo

.....
Approved Signatory

Accredited for compliance with ISO/IEC 17025.
 This report may not be reproduced except in full.



MANUFACTURED AND TESTED BY: Oxoid Australia Pty Limited

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054

Email: oxoid.au@thermofisher.com Website: www.oxoid.com.au