

TaqMan® RNase P Fast 384-Well Instrument Verification Plate

Product No. **4455280**
 Lot No. **1209054**
 Date of Manufacture **30Oct2012**
 Expiration Date **23Jul2013**

TEST	SPECIFICATION	RESULT
The TaqMan® RNase P Fast 384-Well Instrument Verification Plate meets the Applied Biosystems 2-fold Discrimination Test and the Linearity Test specifications when tested on an Applied Biosystems® ViiA™ 7 Real-Time PCR System.		

2-fold Discrimination Test

$(\text{Ave Copy No.})_{10K} - (3 \text{ StdDev})_{10K}$	$(\text{Ave Copy No.})_{10K} - (3\text{StdDev})_{10K} > (\text{Ave Copy No.})_{5K} + (3\text{StdDev})_{5K}$	8483.65 >
$(\text{Ave Copy No.})_{5K} + (3 \text{ StdDev})_{5K}$		5607.48

Linearity Test

R ² of standard curve	R ² > 0.990	0.994
----------------------------------	------------------------	-------

For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.

**ISO9001
REGISTERED**

Singapore

S C Hong

Quality Assurance Manager

Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #05-03,
 Singapore 739256 Tel: (65) 6362 9300

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at cofarequest@appliedbiosystems.com