

CERTIFICATE OF ANALYSIS

PRODUCT PP2423
HAEMOPHILUS ID W/ INDICATOR

LOT NUMBER 1212768

EXPIRY DATE 2012.06.09

RELEASE DATE 2012.04.16

Physical Testing **Results** **Specification**

Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	ORANGE/PINK
Clarity	Pass	Translucent
Gel Strength	Pass	Satisfactory
pH	7.2	7.1 - 7.5
Sterility @ 30°C for 3 days /sample plan	Pass	Within Limits

This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1

Microbiology Testing **Result** **Specification** **Method**

Haemophilus influenzae ATCC® 10211	Pass	Growth around the X & V disc, agar pink	QC34
Haemophilus parainfluenzae ATCC® 7901	Pass	Growth around XV & V disc, agar yellow	QC34

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Paul Costello

Paul Costello

Approved Signatory

Accredited for compliance with ISO/IEC 17025.
This report may not be reproduced except in full.



Accreditation
Number 3186



Accreditation No. 3186

MANUFACTURED AND TESTED BY:

Thermo Fisher Scientific, Microbiology Division

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054

Email: micro_orders@thermofisher.com Website: www.thermofisher.com.au