

Catalog Number L34952
Product Name LIVE/DEAD® FungaLight™ Yeast Viability Kit *for flow cytometry*
Lot Number 1227003

	LOT DATA	SPECIFICATION
ABSORPTION Optical Density of Component A ¹	0.79 at 473 nm	0.75 ± 0.08 at 474 ± 3 nm
ABSORPTION Optical Density of Component B ²	0.61 at 494 nm	0.59 ± 0.06 at 493 ± 3 nm

1. Buffer: phosphate buffered saline, pH 7.2 (301x dilution).
2. Buffer: phosphate buffered saline, pH 7.2 (201x dilution).

Betty Wood

Betty Wood, Quality Assurance Manager
20-Sep-2012

Life Technologies Corporation, on behalf of its Invitrogen business, Molecular Probes® labeling and detection technologies, certifies on the date above that this is an accurate record of the analysis of the subject lot and that the data conform to the specifications in effect for this product at the time of analysis.