



CERTIFICATE OF ANALYSIS

PRODUCT PP2179
MYGP + COPPER

LOT NUMBER 1242832

EXPIRY DATE 2013.02.22

RELEASE DATE 2013.01.03

Physical Testing	Results	Specification
------------------	---------	---------------

Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	Pale Green
Clarity	Pass	Translucent
Gel Strength	Pass	Satisfactory
pH	4.9	4.5 - 4.9
Sterility RT for 2 days/30°C for 5days	Pass	Within Limits

This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1

Microbiology Testing	Result	Specification	Method
----------------------	--------	---------------	--------

Brettanomyces species MVQC 0071	Pass	Growth	QC27
Saccharomyces species B AJL 4053	Pass	No growth	QC27

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Rina Pattullo

.....
Approved Signatory

Accredited for compliance with ISO/IEC 17025.
This report may not be reproduced except in full.



MANUFACTURED AND TESTED BY: Oxoid Australia Pty Limited

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054

Email: oxoid.au@thermofisher.com Website: www.oxoid.com.au