



CERTIFICATE OF ANALYSIS

PRODUCT TM4013
TRIBUTYRIN AGAR 100ML

LOT NUMBER 1258941

EXPIRY DATE 2014.05.21

RELEASE DATE 2013.05.22

Physical Testing **Results** **Specification**
Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	Straw
Clarity	Pass	Translucent
Gel Strength	Pass	Satisfactory
pH	7.6	7.3 - 7.7

Microbiology Testing	Result	Specification	Method
Escherichia coli ATCC® 25922	Pass	White/straw lipolysis-ve cols	QC11
Pseudomonas aeruginosa ATCC® 27853	Pass	White/green lipolysis+ve cols	QC11
Staphylococcus aureus ATCC® 25923	Pass	White/straw lipolysis+ve cols	QC11
Candida albicans ATCC® 10231	Pass	Cream/straw lipolysis+ve cols	QC11
Staphylococcus epidermidis ATCC® 12228	Pass	White lipolysis+ve cols	QC11

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Rina Pattullo

.....
Approved Signatory

Accredited for compliance with ISO/IEC 17025.
 This report may not be reproduced except in full.



MANUFACTURED AND TESTED BY: Oxoid Australia Pty Limited

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054

Email: oxoid.au@thermofisher.com Website: www.oxoid.com.au