

# CERTIFICATE OF ANALYSIS

**PRODUCT** PP2347  
**REINFORCED CLOSTRIDIAL AGAR**

**LOT NUMBER** 1266621

**EXPIRY DATE** 2013.09.21

**RELEASE DATE** 2013.07.29

---

Physical Testing	Results	Specification
------------------	---------	---------------

*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

---

Appearance (Colour)	Pass	Beige / Colourless
Clarity	Pass	Translucent
Gel Strength	Pass	Satisfactory
pH	6.7	6.6 - 7.0
Sterility @ 30°C for 3 days /sample plan	Pass	Within Limits

*This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1*

---

Microbiology Testing	Result	Specification	Method
----------------------	--------	---------------	--------

---

Clostridium perfringens ATCC® 13124	Pass	Colourless colonies	QC27
-------------------------------------	------	---------------------	------

---

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Paul Costello

*Paul Costello*

.....  
**Approved Signatory**

Accredited for compliance with ISO/IEC 17025.  
This report may not be reproduced except in full.



Accreditation  
Number 3186



Accreditation No. 3186

**MANUFACTURED AND TESTED BY:**

**Thermo Fisher Scientific, Microbiology Division**

20 Dalgleish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054

Email: [micro\\_orders@thermofisher.com](mailto:micro_orders@thermofisher.com) Website: [www.thermofisher.com.au](http://www.thermofisher.com.au)