

CERTIFICATE OF ANALYSIS

PRODUCT PP1067P90
CHOCOLATE AGAR + HORSE BLOOD

LOT NUMBER 1275079

EXPIRY DATE 23.01.2013

| Physical Characteristics | Results | Specification |
|--------------------------|---------------------|---------------------|
| Appearance | Chocolate brown gel | Chocolate brown gel |
| pH | 7.24 | 7.00 - 7.40 |

Sterility: Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

| Microbiological Characteristics | Results | Specification Quantitative Tests |
|--|---------------------------|----------------------------------|
| CO₂, 35°C, 24-48hr | | |
| <i>Neisseria gonorrhoeae</i> ATCC®43069 | 1-2mm colourless colonies | |
| <i>Haemophilus influenzae</i> ATCC®10211 | 2-3mm colourless colonies | |

Released By



Zulhusni Amir

MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)

(Formerly known as Biomedica Laboratories Sdn Bhd)

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: oxid.myqcbml@thermofisher.com