



CERTIFICATE OF ANALYSIS

PRODUCT TM0098
MANNITOL SELENITE 10ML

LOT NUMBER 1277914

EXPIRY DATE 2014.02.28

RELEASE DATE 2013.10.31

Physical Testing **Results** **Specification**
Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	C/less to Pale orange, mi
Clarity	Pass	Translucent
pH	7.1	6.9 - 7.3

Microbiological Testing	Result	Specification	Method
Escherichia coli ATCC® 25922	Pass	Inhibited	QC17
Salmonella typhimurium ATCC® 14028	Pass	Growth	QC17
Shigella sonnei ATCC® 9290	Pass	Growth	QC17
Escherichia coli/Salmonella typhimurium	Pass	E.coli Inhibited/S.typh Growth	QC17

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

The product is stable to the stated expiry date when stored in accordance with instructions per product label.

Released by: Rina Pattullo

.....
Approved Signatory

Accredited for compliance with ISO/IEC 17025.
 This report may not be reproduced except in full.



MANUFACTURED AND TESTED BY: Oxoid Australia Pty Limited

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054

Email: oxoid.au@thermofisher.com Website: www.oxoid.com.au