



CERTIFICATE OF ANALYSIS

PRODUCT TM4068
THIOGLYCOLLATE BROTH 5ML (PK 20)

LOT NUMBER 1281073

EXPIRY DATE 2014.12.02

RELEASE DATE 2013.12.16

Physical Testing	Results	Specification
-------------------------	----------------	----------------------

Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	Yellow
Clarity	Pass	Translucent
pH	7.1	6.9 - 7.3
Sterility @ 35°C for 14days /sample plan	Pass	Within Limits
<i>This test is performed in accordance Oxoid method QC21, NATA FADs, ASM guidelines: Normal inspection AQL 2.5, General inspection level=1</i>		
Sterility @ RT for 14 days /sample plan	Pass	Within Limits
<i>This test is performed in accordance Oxoid method QC21, NATA FADs, ASM guidelines: Normal inspection AQL 2.5, General inspection level=1</i>		

Microbiological Testing	Result	Specification	Method
Clostridium sporogenes ATCC® 19404	Pass	Growth	QC27
Staphylococcus aureus ATCC® 6538	Pass	Growth	QC27
Pseudomonas aeruginosa ATCC® 9027	Pass	Growth	QC27

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

The product is stable to the stated expiry date when stored in accordance with instructions per product label.

Released by: Paul Costello

Approved Signatory

Accredited for compliance with ISO/IEC 17025.
This report may not be reproduced except in full.



MANUFACTURED AND TESTED BY: Oxoid Australia Pty Limited
 20 Dalglish Street, Thebarton, South Australia 5031
 Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054
 Email: oxoid.au@thermofisher.com Website: www.oxoid.com.au